NOTES:

- 1. SUBSTRATE FUSED SILICA
- 2. SURFACE S2 TO BE PARALLEL TO SURFACE S1 TO WITHIN <3 ARCMIN
- 3. COATING (APPLY ACROSS COATING APERTURE):

S1: 266 HR Coating R (ABS) > 99.80% @ 266nm @ 0-45° AOI


DAMAGE THRESHOLD, PULSED: 2.0 J/cm², 20ns, 20Hz @ 266nm CW: 1MW/cm² @ 266nm


S2: NONE

4. FINE GRIND SURFACE

- POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. CLEAR APERTURE AND COATING APERTURE ARE CENTERED ON SURFACES

APPLY ARROW ON EDGE WITH LASER ETCH, PENCIL, OR PERMANENT INK POINTING TOWARDS SURFACE ST





FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

| | S1 | S2 | | | (R) | |
|----------------------|----------------------|----------------------|------------------------|--------|--------------------------------------|----|
| SHAPE | PLANO | PLANO | | | Edmund Optics® | 3) |
| SURFACE QUALITY | 10-5 | COMMERCIAL POLISH | | | | |
| SURFACE FLATNESS | 0.10 WAVE | N/A | | | Ø12.7mm x 6.35mm,266nm, NdYAG MIRROI |)R |
| MIN CLEAR APERTURE | Ø11.43 | N/A | THIRD ANGLE PROJECTION | TITLE | 0-45° AOI | 1 |
| MIN COATING APERTURE | Ø11.43 | N/A | | | | |
| BEVEL | PROTECTIVE AS NEEDED | PROTECTIVE AS NEEDED | ALL DIMS IN mm | DWG NO | 38833 SHE 1 O | |