## NOTES:


- SUBSTRATE: II-VI Infrared ZnSe
- 2. CENTERING TOLERANCE: EDGE THICKNESS VARIATION MEASURED AT THE CLEAR APERTURE OF \$1 NOT TO EXCEED 12.7µm
- 3. COATING (APPLY ACROSS COATING APERTURE)

S1 & S2: NONE

4.\ FINE GROUND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. WAVELENGTH RANGE: 600 18000nm
- 7. SURFACE ROUGHNESS: <50 Å
- 8. ROHS: COMPLIANT





**SECTION A-A** 

	\$1	\$2	
SHAPE	CONVEX	PLANO	
RADIUS	210.40	INFINITY	
SURFACE QUALITY	40-20	40-20	
CLEAR APERTURE	90%	90%	
POWER at 10.6µm	λ/10	λ/10	
IRREGULARITY at 10.6µm	λ/20	λ/20	
BEVEL	PROTECTIVE AS NEEDED PROTECTIVE AS NEE		

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

1 OF 1

EFL (AT 10.6 microns) BFL (AT 10.6 microns)	150.00 148.88		<b>Edmund Optic</b>	S®
THIRD ANGLE PROJECTION	$\phi$	TITLE	25.4mm Dia. x 150mm FL, Uncoated, ZnSe Plano-Convex Lens	
ALL DIMS IN	mm	DWG NO	39535	SHEET

39535