

## NOTES:


- 1. SUBSTRATE: N-BK7
- 2. S2 TO BE PARALLEL WITH \$1 TO <1 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE) S1 & S2: R(ABS) <0.25% @ 532nm R(ABS) < 0.25% @ 1064nm R(AVG) <1.0% @ 500 - 1100nm



- 5. Rohs Compliant
- 6. SURFACE FLATNESS (P-V):  $\lambda/4$  (per inch over clear aperture)







## **PARTS TO THIS DRAWING**

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	60-40	60-40
CLEAR APERTURE	11.25 x 11.25	11.25 x 11.25
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

			<b>Edmund Optics®</b>		;S®
	THIRD ANGLE _ PROJECTION		TITLE	12.5mm SQ., 3mm THICK, YAG-BBAR COATED, λ/4 N-BK7 WINDOW	
_	ALL DIMS IN	mm	DWG NO	13313	SHEET

TITLE	12.5mm SQ., 3mm THICK, YAG-BBA COATED, λ/4 N-BK7 WINDOW	R
DWG NO	13313	SHEET

1 OF 1