NOTES:


1. SUBSTRATE: N-BK7


- 2. S2 TO BE PARALLEL WITH S1 TO <1 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE) S1 & S2: R(ABS) <0.25% @ 532nm R(ABS) < 0.25% @ 1064nm R(AVG) <1.0% @ 500 - 1100nm




5. Rohs Compliant

6. SURFACE FLATNESS (P-V): $\lambda/4$ (per inch over clear aperture)







PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2
SHAPE	PLANO	PLANO
SURFACE QUALITY	60-40	60-40
CLEAR APERTURE	67.50 x 67.50	67.50 x 67.50
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

			Edmund Optics®	
THIRD ANGLE PROJECTION	\$	TITLE	75mm SQ., 4mm THICK, YAG-BBAR CO λ/4 N-BK7 WINDOW	ATED,
ALL DIMS IN	mm	DWG NO	15798	SHEET 1 OF 1